

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051821\  
 Data File : VN067120.D  
 Acq On : 18 May 2021 12:00  
 Operator : JC/MD  
 Sample : VN0518WBL01  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0518WBL01

Quant Time: May 19 02:46:50 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051421W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue May 18 04:00:15 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 7.600  | 168   | 620776   | 50.00    | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.525  | 114   | 863655   | 50.00    | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.360 | 117   | 742777   | 50.00    | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.304 | 152   | 282090   | 50.00    | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 7.962  | 65    | 284855   | 42.48    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =     | 84.96%   |
| 35) Dibromofluoromethane    | 7.519  | 113   | 261532   | 46.12    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =     | 92.24%   |
| 50) Toluene-d8              | 10.037 | 98    | 1024129  | 47.65    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =     | 95.30%   |
| 62) 4-Bromofluorobenzene    | 12.360 | 95    | 308158   | 42.80    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =     | 85.60%   |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051821\  
Data File : VN067120.D  
Acq On : 18 May 2021 12:00  
Operator : JC/MD  
Sample : VN0518WBL01  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0518WBL01

Quant Time: May 19 02:46:50 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051421W.M  
Quant Title : SW846 8260  
QLast Update : Tue May 18 04:00:15 2021  
Response via : Initial Calibration

