

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051821\
 Data File : VN067129.D
 Acq On : 18 May 2021 16:03
 Operator : JC/MD
 Sample : M2387-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41801

Quant Time: May 19 02:50:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.597	168	649247	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	901528	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	767296	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.301	152	325210	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.959	65	317258	45.23	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	90.46%		
35) Dibromofluoromethane	7.519	113	292658	49.44	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.88%		
50) Toluene-d8	10.037	98	1109949	49.47	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.94%		
62) 4-Bromofluorobenzene	12.357	95	332267	44.21	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.42%		
Target Compounds						Qvalue
11) Tert butyl alcohol	4.724	59	51297	43.87	ug/l #	77
16) Acetone	3.772	43	45398	20.83	ug/l	98
18) Methyl Acetate	4.274	43	13358	2.06	ug/l #	74
25) 2-Butanone	6.760	43	48838	14.37	ug/l	98
40) Benzene	7.975	78	77330	3.10	ug/l	99
52) Toluene	10.102	92	72303	4.48	ug/l	99
67) Ethyl Benzene	11.470	91	11568	0.40	ug/l	98
68) m/p-Xylenes	11.574	106	25625	2.30	ug/l	89
69) o-Xylene	11.907	106	10269	0.92	ug/l	96
80) 1,3,5-Trimethylbenzene	12.693	105	8545	0.38	ug/l	90
84) 1,2,4-Trimethylbenzene	13.001	105	23164	1.11	ug/l	100
95) Naphthalene	15.098	128	18453	4.50	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051821\
Data File : VN067129.D
Acq On : 18 May 2021 16:03
Operator : JC/MD
Sample : M2387-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41801

Quant Time: May 19 02:50:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

