

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
 Data File : VN077737.D
 Acq On : 18 May 2023 15:32
 Operator : JC\MD
 Sample : 02834-08 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D1-20230516

Quant Time: May 19 07:35:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/19/2023
 Supervised By : Mahesh Dadoda 05/19/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	369498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	662791	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	584692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	213967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	253857	45.917	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.840%		
35) Dibromofluoromethane	8.177	113	214090	49.535	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.080%		
50) Toluene-d8	10.571	98	818752	49.217	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.440%		
62) 4-Bromofluorobenzene	12.859	95	272013	46.297	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.600%		
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.377	101	5848m	1.430	ug/l	
12) 1,1-Dichloroethene	4.347	96	3334	0.858	ug/l #	93
27) cis-1,2-Dichloroethene	7.488	96	2564	0.490	ug/l	62
44) Trichloroethene	9.359	130	329840	69.829	ug/l	100
64) Tetrachloroethene	11.112	164	2596	0.635	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
Data File : VN077737.D
Acq On : 18 May 2023 15:32
Operator : JC\MD
Sample : 02834-08 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D1-20230516

Quant Time: May 19 07:35:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/19/2023
Supervised By : Mahesh Dadoda 05/19/2023

