

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
 Data File : VN077739.D
 Acq On : 18 May 2023 16:19
 Operator : JC\MD
 Sample : 02834-06 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D2-20230516

Quant Time: May 19 07:35:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	366348	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	668533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	599474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	224288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	250762	45.747	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.500%
35) Dibromofluoromethane	8.177	113	210500	48.286	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	10.577	98	830566	49.498	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.000%
62) 4-Bromofluorobenzene	12.853	95	281954	47.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.160%
Target Compounds						
44) Trichloroethene	9.359	130	432658	90.809	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
Data File : VN077739.D
Acq On : 18 May 2023 16:19
Operator : JC\MD
Sample : 02834-06 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D2-20230516

Quant Time: May 19 07:35:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

