

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
 Data File : VN077741.D
 Acq On : 18 May 2023 17:06
 Operator : JC\MD
 Sample : 02834-14DL 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP02-20230517DL

Quant Time: May 19 07:36:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	356401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	647144	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	551279	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	180721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	238849	44.790	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.580%
35) Dibromofluoromethane	8.171	113	204169	48.382	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.760%
50) Toluene-d8	10.576	98	794431	48.909	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.820%
62) 4-Bromofluorobenzene	12.859	95	244494	42.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.240%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.389	101	12893	3.269	ug/l	95
44) Trichloroethene	9.359	130	13345	2.894	ug/l	89
64) Tetrachloroethene	11.112	164	1164	0.302	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
Data File : VN077741.D
Acq On : 18 May 2023 17:06
Operator : JC\MD
Sample : 02834-14DL 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP02-20230517DL

Quant Time: May 19 07:36:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

