

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
 Data File : VN077744.D
 Acq On : 18 May 2023 18:17
 Operator : JC\MD
 Sample : 02834-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D3-20230516

Quant Time: May 19 07:36:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	348623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	646554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	577257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	202124	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	245776	47.117	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.240%		
35) Dibromofluoromethane	8.177	113	202009	47.914	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.820%		
50) Toluene-d8	10.577	98	797970	49.172	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.340%		
62) 4-Bromofluorobenzene	12.859	95	260043	45.371	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.740%		
Target Compounds						
44) Trichloroethene	9.359	130	10155	2.204	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
Data File : VN077744.D
Acq On : 18 May 2023 18:17
Operator : JC\MD
Sample : 02834-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D3-20230516

Quant Time: May 19 07:36:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

