

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
 Data File : VN077745.D
 Acq On : 18 May 2023 18:40
 Operator : JC\MD
 Sample : 02834-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D1-20230516

Quant Time: May 19 07:37:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	344003	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	627986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	544441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	181293	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	237958	46.231	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.460%
35) Dibromofluoromethane	8.177	113	199570	48.735	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	10.577	98	791835	50.237	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.480%
62) 4-Bromofluorobenzene	12.853	95	239712	43.061	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.120%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.395	101	8552	2.246	ug/l	99
16) Acetone	4.459	43	4522	2.792	ug/l	94
27) cis-1,2-Dichloroethene	7.500	96	8240	1.691	ug/l	88
44) Trichloroethene	9.359	130	1408423	314.696	ug/l	100
64) Tetrachloroethene	11.112	164	8236	2.163	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
Data File : VN077745.D
Acq On : 18 May 2023 18:40
Operator : JC\MD
Sample : 02834-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D1-20230516

Quant Time: May 19 07:37:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

