

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
 Data File : VN077747.D
 Acq On : 18 May 2023 19:27
 Operator : JC\MD
 Sample : 02834-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D1-20230517

Quant Time: May 19 07:37:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	357667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	648642	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	564819	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	183601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	242882	45.385	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.760%		
35) Dibromofluoromethane	8.177	113	205098	48.490	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.980%		
50) Toluene-d8	10.576	98	811460	49.842	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.680%		
62) 4-Bromofluorobenzene	12.853	95	249518	43.395	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.780%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.383	101	13685	3.458	ug/l #	88
12) 1,1-Dichloroethene	4.359	96	4426	1.177	ug/l #	83
19) Methyl tert-butyl Ether	5.812	73	10457	0.768	ug/l	92
27) cis-1,2-Dichloroethene	7.488	96	14233	2.809	ug/l	92
30) Chloroform	7.977	83	5623	0.644	ug/l	100
44) Trichloroethene	9.359	130	874677	189.213	ug/l	97
49) 1,4-Dioxane	9.706	88	1490	18.474	ug/l #	74
64) Tetrachloroethene	11.112	164	41244	10.442	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051823\
Data File : VN077747.D
Acq On : 18 May 2023 19:27
Operator : JC\MD
Sample : 02834-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE125D1-20230517

Quant Time: May 19 07:37:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

