

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051920\
 Data File : VN061490.D
 Acq On : 19 May 2020 17:53
 Operator : JC/MD
 Sample : PB128977TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: May 20 02:59:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	290988	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	530928	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	508205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	211871	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	201348	46.36	ug/l	0.00
Spiked Amount			Recovery	=	92.72%	
35) Dibromofluoromethane	7.55	113	149236	53.66	ug/l	0.00
Spiked Amount			Recovery	=	107.32%	
50) Toluene-d8	10.06	98	658131	51.02	ug/l	0.00
Spiked Amount			Recovery	=	102.04%	
62) 4-Bromofluorobenzene	12.37	95	243103	50.44	ug/l	0.00
Spiked Amount			Recovery	=	100.88%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.55	94	251	Below Cal	#	39
16) Acetone	3.77	43	13429	0.676	ug/l	92
17) Carbon Disulfide	4.01	76	474	Below Cal	#	75
18) Methyl Acetate	4.27	43	1236	Below Cal	#	81
20) Methylene Chloride	4.50	84	13017	3.647	ug/l	94
25) 2-Butanone	6.79	43	1824	0.922	ug/l	90
29) Tetrahydrofuran	7.17	42	345	0.263	ug/l #	44
31) Cyclohexane	7.62	56	6403	Below Cal	#	2
36) 1,1-Dichloropropene	7.63	75	23697	4.773	ug/l #	50
38) Carbon Tetrachloride	7.62	117	29352	6.421	ug/l #	16
43) Isopropyl Acetate	8.13	43	90513	11.527	ug/l #	90
45) 1,2-Dichloropropane	9.24	63	1838	0.491	ug/l #	43
48) Methyl methacrylate	9.15	41	3751	1.034	ug/l #	9
49) 1,4-Dioxane	9.18	88	61	Below Cal	#	24
51) 4-Methyl-2-Pentanone	9.95	43	7419	1.734	ug/l	97
54) cis-1,3-Dichloropropene	9.87	75	47786	7.415	ug/l #	62
57) 1,3-Dichloropropane	10.74	76	7355	1.172	ug/l #	42
59) 2-Hexanone	10.74	43	90607	29.277	ug/l #	49
68) m/p-Xylenes	11.59	106	2713	0.378	ug/l	98
71) Bromoform	12.08	173	61	2.040	ug/l #	29
74) N-amyl acetate	12.00	43	4454	0.715	ug/l #	55
76) 1,2,3-Trichloropropane	12.37	75	125362	27.575	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.71	105	9217	0.662	ug/l	91
81) trans-1,4-Dichloro-2-buten	12.37	75	125362	73.150	ug/l #	12
84) 1,2,4-Trimethylbenzene	13.01	105	2989	0.214	ug/l	94
95) Naphthalene	15.10	128	3684	0.330	ug/l #	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051920\
Data File : VN061490.D
Acq On : 19 May 2020 17:53
Operator : JC/MD
Sample : PB128977TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: May 20 02:59:17 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. 0.00 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 290988

Ion Ratio Lower Upper

168 100

99 62.0 51.1 76.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



7.63

100000

50000

0

Time--> 7.50 7.55 7.60 7.65 7.70



#5

Bromomethane

Concen: Below Cal

RT: 2.55 min Scan# 241

Delta R.T. 0.01 min

Lab File: VN061490.D


Acq: 19 May 2020 17:53

Tgt Ion: 94 Resp: 251

Ion Ratio Lower Upper

94 100

96 35.2 74.9 112.3#



Abundance Ion 93.90 (93.60 to 94.60): VNO

Ion 95.90 (95.60 to 96.60): VNO

500

400


300


200

100

0

Time--> 2.52 2.54 2.56





#16

Acetone

Concen: 0.676 ug/l

RT: 3.77 min Scan# 622

Delta R.T. 0.00 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 13429

Ion Ratio Lower Upper

43 100

58 27.5 25.6 38.4



Abundance Ion 43.00 (42.70 to 43.70): VN061429.D (-589) (-)

Ion 58.00 (57.70 to 58.70): VN061429.D (-589) (-)

3.77

6000

5000

4000

3000

2000

1000

0



Time-->

3.70

3.75

3.80

3.85



#17

Carbon Disulfide

Concen: Below Cal

RT: 4.01 min Scan# 695

Delta R.T. -0.00 min

Lab File: VN061490.D


Acq: 19 May 2020 17:53

Tgt Ion: 76 Resp: 474

Ion Ratio Lower Upper

76 100

78 0.0 7.4 11.0#



Abundance Ion 75.90 (75.60 to 76.60): VN061429.D (-675) (-)

Ion 77.90 (77.60 to 78.60): VN061429.D (-675) (-)

4.01

300

200

100

0


Time-->


3.98

4.00

4.02

4.04









#18
Methyl Acetate
Concen: Below Cal
RT: 4.27 min Scan# 777
Delta R.T. -0.01 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 1236
Ion Ratio Lower Upper
43 100
74 35.9 20.8 31.2#



#20
Methylene Chloride
Concen: 3.647 ug/l
RT: 4.50 min Scan# 849
Delta R.T. 0.00 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53

Tgt Ion: 84 Resp: 13017
Ion Ratio Lower Upper
84 100
49 111.8 95.3 142.9
51 40.0 29.0 43.6
86 66.1 50.2 75.2









#25
2-Butanone
Concen: 0.922 ug/l
RT: 6.79 min Scan# 1559
Delta R.T. 0.00 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 1824
Ion Ratio Lower Upper
43 100
72 23.3 23.0 34.6



#29
Tetrahydrofuran
Concen: 0.263 ug/l
RT: 7.17 min Scan# 1677
Delta R.T. 0.00 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53

Tgt Ion: 42 Resp: 345
Ion Ratio Lower Upper
42 100
72 19.1 39.0 58.6#
71 0.0 36.9 55.3#









#31
Cyclohexane
Concen: Below Cal
RT: 7.62 min Scan# 1818
Delta R.T. 0.01 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 56 Resp: 6403
Ion Ratio Lower Upper
56 100
69 173.3 27.8 41.6#
84 132.3 72.7 109.1#




Abundance Ion 56.00 (55.70 to 56.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 84.00 (83.70 to 84.70): VNO






#33
1,2-Dichloroethane-d4
Concen: 46.363 ug/l
RT: 7.99 min Scan# 1933
Delta R.T. 0.00 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53

Tgt Ion: 65 Resp: 201348
Ion Ratio Lower Upper
65 100
67 52.3 0.0 101.4



Abundance Ion 64.90 (64.60 to 65.60): VNO
Ion 66.90 (66.60 to 67.60): VNO





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2108

Delta R.T. 0.00 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53

Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:114 Resp: 530928

Ion Ratio Lower Upper

114 100

63 19.4 0.0 40.8




88 15.7 0.0 30.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 53.655 ug/l

RT: 7.55 min Scan# 1796

Delta R.T. 0.00 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53


Tgt Ion:113 Resp: 149236

Ion Ratio Lower Upper

113 100

111 101.7 80.6 121.0



192 17.5 13.2 19.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.773 ug/l
 RT: 7.63 min Scan# 1820
 Delta R.T. -0.13 min
 Lab File: VN061490.D
 Acq: 19 May 2020 17:53




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion	75	Resp	23697
Ion Ratio	Lower	Upper	
75	100		
110	9.1	17.3	51.9#
77	0.0	24.9	37.3#



#38
 Carbon Tetrachloride
 Concen: 6.421 ug/l
 RT: 7.62 min Scan# 1819
 Delta R.T. -0.11 min
 Lab File: VN061490.D
 Acq: 19 May 2020 17:53

Tgt Ion	117	Resp	29352
Ion Ratio <td>Lower</td> <td>Upper</td> <td></td>	Lower	Upper	
117	100		
119	4.9	77.0	115.6#
121	0.0	24.7	37.1#





#43

Isopropyl Acetate

Concen: 11.527 ug/l

RT: 8.13 min Scan# 1976

Delta R.T. 0.00 min

Lab File: VN061490.D





Acq: 19 May 2020 17:53

Instrument :

MSVOA_N

ClientSampleId :

Tot Ion:	43	Resp:	90513
Ion	Ratio	Lower	Upper
43	100		
61	21.4	23.0	34.4#
87	15.1	11.6	17.4



#45

1,2-Dichloropropane

Concen: 0.491 ug/l




RT: 9.24 min Scan# 2321


Delta R.T. 0.15 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53

Tgt Ion:	63	Resp:	1838
Ion	Ratio	Lower	Upper
63	100		
65	0.0	25.2	37.8#









#48
Methvl methacrylate
Concen: 1.034 ug/l
RT: 9.15 min Scan# 2293
Delta R.T. -0.02 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 41 Resp: 3751
Ion Ratio Lower Upper
41 100
69 0.0 75.3 112.9#
39 0.0 52.6 79.0#



#49
1,4-Dioxane
Concen: Below Cal
RT: 9.18 min Scan# 2302
Delta R.T. 0.01 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53

Tgt Ion: 88 Resp: 61
Ion Ratio Lower Upper
88 100
43 0.0 26.3 39.5#
58 0.0 54.1 81.1#





#50

Toluene-d8

Concen: 51.015 ug/l

RT: 10.06 min Scan# 2577

Delta R.T. 0.00 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 98 Resp: 658131

Ion Ratio Lower Upper

98 100

100 65.6 51.8 77.8



#51

4-Methyl-2-Pentanone

Concen: 1.734 ug/l

RT: 9.95 min Scan# 2544

Delta R.T. 0.00 min

Lab File: VN061490.D




Acq: 19 May 2020 17:53


Tgt Ion: 43 Resp: 7419

Ion Ratio Lower Upper

43 100

58 43.8 33.4 50.2









#54
 cis-1,3-Dichloropropene
 Concen: 7.415 ug/l
 RT: 9.87 min Scan# 2519
 Delta R.T. 0.06 min
 Lab File: VN061490.D
 Acq: 19 May 2020 17:53




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 75 Resp: 47786
 Ion Ratio Lower Upper
 75 100
 77 1.0 24.8 37.2#
 39 30.3 39.8 59.6#



#57
 1,3-Dichloropropane
 Concen: 1.172 ug/l
 RT: 10.74 min Scan# 2788
 Delta R.T. 0.06 min
 Lab File: VN061490.D
 Acq: 19 May 2020 17:53

Tgt Ion: 76 Resp: 7355
 Ion Ratio Lower Upper
 76 100
 78 0.0 26.0 39.0#









#59
2-Hexanone
Concen: 29.277 ug/l
RT: 10.74 min Scan# 2788
Delta R.T. 0.02 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 43 Resp: 90607
Ion Ratio Lower Upper
43 100
58 20.3 29.2 87.6#



#62
4-Bromofluorobenzene
Concen: 50.438 ug/l
RT: 12.37 min Scan# 3297
Delta R.T. 0.00 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53

Tgt Ion: 95 Resp: 243103
Ion Ratio Lower Upper
95 100
174 72.1 0.0 141.2
176 69.9 0.0 137.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. 0.00 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
117	100		
82	55.3	45.2	67.8
119	31.1	25.3	37.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VN
Ion 118.90 (118.60 to 119.60): V






#68
m/p-Xylenes
Concen: 0.378 ug/l
RT: 11.59 min Scan# 3053
Delta R.T. -0.00 min
Lab File: VN061490.D
Acq: 19 May 2020 17:53

Tgt Ion	Ratio	Lower	Upper
106	100		
91	206.7	163.0	244.6



Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VN





#71

Bromoform

Concen: 2.040 ug/l

RT: 12.08 min Scan# 3206

Delta R.T. -0.02 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:173 Resp: 61

Ion Ratio Lower Upper

173 100

175 0.0 23.9 71.8#




254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. 0.00 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53


Tgt Ion:152 Resp: 211871

Ion Ratio Lower Upper

152 100

115 61.0 30.9 92.5



150 159.2 0.0 349.2




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 0.715 ug/l

RT: 12.00 min Scan# 3182

Delta R.T. -0.04 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53

Instrument :

MSVOA_N

ClientSampleId :

Tot Ion: 43 Resp: 4454


Ion Ratio Lower Upper

43 100

70 10.2 37.0 55.4#

55 14.9 19.5 29.3#

61 0.0 20.8 31.2#






Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0



#76

1,2,3-Trichloropropane

Concen: 27.575 ug/l

RT: 12.37 min Scan# 3297

Delta R.T. -0.15 min

Lab File: VN061490.D


Acq: 19 May 2020 17:53

Tgt Ion: 75 Resp: 125362

Ion Ratio Lower Upper



75 100


77 0.9 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0





#80

1,3,5-Trimethylbenzene

Concen: 0.662 ug/l

RT: 12.71 min Scan# 3401

Delta R.T. 0.00 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion:105 Resp: 9217

Ion Ratio Lower Upper

105 100

120 54.6 24.1 72.4



#81

trans-1,4-Dichloro-2-butene

Concen: 73.150 ug/l

RT: 12.37 min Scan# 3297

Delta R.T. 0.10 min

Lab File: VN061490.D

Acq: 19 May 2020 17:53




Tgt Ion: 75 Resp: 125362


Ion Ratio Lower Upper

75 100

53 0.1 75.2 112.8#

89 0.1 38.2 57.4#









#84
 1,2,4-Trimethylbenzene
 Concen: 0.214 ug/l
 RT: 13.01 min Scan# 3495
 Delta R.T. -0.00 min
 Lab File: VN061490.D
 Acq: 19 May 2020 17:53

Instrument :
 MSVOA_N
 ClientSampled :

Tot Ion:105 Resp: 2989
 Ion Ratio Lower Upper
 105 100
 120 48.7 22.4 67.3



#95
 Naphthalene
 Concen: 0.330 ug/l
 RT: 15.10 min Scan# 4146
 Delta R.T. 0.00 min
 Lab File: VN061490.D
 Acq: 19 May 2020 17:53

Tgt Ion:128 Resp: 3684
 Ion Ratio Lower Upper
 128 100
 127 16.9 10.1 15.1#
 129 12.3 8.6 13.0

