

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051920\
 Data File : VN061476.D
 Acq On : 19 May 2020 12:05
 Operator : JC/MD
 Sample : VN0519WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0519WBL01

Quant Time: May 20 04:10:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	242915	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	448220	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	420258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	166435	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	170938	47.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.30%	
35) Dibromofluoromethane	7.54	113	125813	53.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.16%	
50) Toluene-d8	10.06	98	553919	50.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.72%	
62) 4-Bromofluorobenzene	12.37	95	197646	48.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051920\
Data File : VN061476.D
Acq On : 19 May 2020 12:05
Operator : JC/MD
Sample : VN0519WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0519WBL01

Quant Time: May 20 04:10:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

