

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051920\
 Data File : VN061492.D
 Acq On : 19 May 2020 18:40
 Operator : JC/MD
 Sample : L2676-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S32-0690

Quant Time: May 20 06:14:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	285102	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	524248	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	497292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	203576	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	199548	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
35) Dibromofluoromethane	7.55	113	152774	55.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.26%	
50) Toluene-d8	10.06	98	647726	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	12.37	95	236219	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.83	85	4072	1.718	ug/l	99
4) Vinyl Chloride	2.16	62	236081	42.662	ug/l	99
12) 1,1-Dichloroethene	3.70	96	3677	1.258	ug/l	86
21) trans-1,2-Dichloroethene	4.99	96	11572	3.434	ug/l	96
27) cis-1,2-Dichloroethene	6.78	96	185288	46.549	ug/l	85
44) Trichloroethene	8.80	130	24213	6.255	ug/l	99
68) m/p-Xylenes	11.60	106	2195	0.312	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051920\
Data File : VN061492.D
Acq On : 19 May 2020 18:40
Operator : JC/MD
Sample : L2676-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S32-0690

Quant Time: May 20 06:14:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

