

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051921\
 Data File : VN067150.D
 Acq On : 19 May 2021 14:43
 Operator : JC/MD
 Sample : M2420-06ME
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PLAYGROUND-CP-1ME

Manual Integrations
 APPROVED

SAM
 5/20/2021 6:04:15 PM

Quant Time: May 20 01:55:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.589	168	544838	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.522	114	755793	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	646567	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.304	152	274584	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	260160	44.20	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	88.40%		
35) Dibromofluoromethane	7.516	113	246517	49.67	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.34%		
50) Toluene-d8	10.035	98	925521	49.21	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.42%		
62) 4-Bromofluorobenzene	12.363	95	300901m	47.76	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	95.52%		
Target Compounds					Qvalue	
52) Toluene	10.102	92	71251	5.27	ug/l	98
65) Chlorobenzene	11.384	112	889090	61.66	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051921\
Data File : VN067150.D
Acq On : 19 May 2021 14:43
Operator : JC/MD
Sample : M2420-06ME
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PLAYGROUND-CP-1ME

Manual Integrations
APPROVED

SAM
5/20/2021 6:04:15 PM

Quant Time: May 20 01:55:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

