

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051921\  
 Data File : VN067151.D  
 Acq On : 19 May 2021 15:07  
 Operator : JC/MD  
 Sample : IBLK  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 IBLK

Quant Time: May 20 01:55:52 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051421W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue May 18 04:00:15 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 7.597  | 168   | 652683   | 50.00    | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.525  | 114   | 894787   | 50.00    | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.360 | 117   | 768950   | 50.00    | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.304 | 152   | 258785   | 50.00    | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 7.962  | 65    | 302556   | 42.91    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =     | 85.82%   |
| 35) Dibromofluoromethane    | 7.522  | 113   | 282358   | 48.06    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =     | 96.12%   |
| 50) Toluene-d8              | 10.038 | 98    | 1103748  | 49.57    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =     | 99.14%   |
| 62) 4-Bromofluorobenzene    | 12.360 | 95    | 308401   | 41.35    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =     | 82.70%   |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051921\  
Data File : VN067151.D  
Acq On : 19 May 2021 15:07  
Operator : JC/MD  
Sample : IBLK  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
IBLK

Quant Time: May 20 01:55:52 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051421W.M  
Quant Title : SW846 8260  
QLast Update : Tue May 18 04:00:15 2021  
Response via : Initial Calibration

