

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051921\
 Data File : VN067152.D
 Acq On : 19 May 2021 15:31
 Operator : JC/MD
 Sample : M2366-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JG-SPOILS-5-11-21-ARE

Quant Time: May 20 01:56:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 18 04:00:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.600	168	653916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.525	114	897361	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.360	117	744842	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.304	152	300945	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.962	65	311927	44.15	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	88.30%		
35) Dibromofluoromethane	7.522	113	197353	33.49	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	66.98%#		
50) Toluene-d8	10.037	98	1136796	50.90	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.80%		
62) 4-Bromofluorobenzene	12.360	95	337498	45.12	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.24%		
Target Compounds					Qvalue	
16) Acetone	3.770	43	124802	56.84	ug/l	95
20) Methylene Chloride	4.488	84	33026	4.67	ug/l	96
25) 2-Butanone	6.763	43	34885	10.19	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051921\
Data File : VN067152.D
Acq On : 19 May 2021 15:31
Operator : JC/MD
Sample : M2366-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JG-SPOILS-5-11-21-ARE

Quant Time: May 20 01:56:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051421W.M
Quant Title : SW846 8260
QLast Update : Tue May 18 04:00:15 2021
Response via : Initial Calibration

