

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051922\  
 Data File : VN072551.D  
 Acq On : 19 May 2022 13:18  
 Operator : JC\MD  
 Sample : VN0519WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0519WBL01

Quant Time: May 20 06:51:06 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051822W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 19 02:22:47 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 8.081  | 168   | 297777   | 50.000   | ug/l       | # 0.00   |
| 34) 1,4-Difluorobenzene     | 8.963  | 114   | 452075   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 11.739 | 117   | 395133   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.668 | 152   | 150127   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 8.434  | 65    | 251088   | 54.270   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | = 108.540% |          |
| 35) Dibromofluoromethane    | 8.016  | 113   | 170944   | 56.042   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | = 112.080% |          |
| 50) Toluene-d8              | 10.439 | 98    | 613653   | 53.148   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | = 106.300% |          |
| 62) 4-Bromofluorobenzene    | 12.727 | 95    | 195757   | 44.634   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | = 89.260%  |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051922\  
Data File : VN072551.D  
Acq On : 19 May 2022 13:18  
Operator : JC\MD  
Sample : VN0519WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0519WBL01

Quant Time: May 20 06:51:06 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051822W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 19 02:22:47 2022  
Response via : Initial Calibration

