

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
 Data File : VN072558.D
 Acq On : 19 May 2022 16:26
 Operator : JC\MD
 Sample : N2939-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B2-WATER

Quant Time: May 20 06:54:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 02:22:47 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/20/2022
 Supervised By : Mahesh Dadoda 05/20/2022

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	281171	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	412059	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	364948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	133628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	240851	55.132	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	110.260%	
35) Dibromofluoromethane	8.016	113	159267	57.284	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	114.560%	
50) Toluene-d8	10.439	98	569903	54.153	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	108.300%	
62) 4-Bromofluorobenzene	12.727	95	185241	46.338	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	92.680%	
Target Compounds						
16) Acetone	4.298	43	6494m	4.904	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
Data File : VN072558.D
Acq On : 19 May 2022 16:26
Operator : JC\MD
Sample : N2939-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B2-WATER

Quant Time: May 20 06:54:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
Quant Title : SW846 8260
QLast Update : Thu May 19 02:22:47 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/20/2022
Supervised By :Mahesh Dadoda 05/20/2022

