

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
 Data File : VN072560.D
 Acq On : 19 May 2022 17:15
 Operator : JC\MD
 Sample : N2965-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30985

Quant Time: May 20 06:54:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 02:22:47 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	277650	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	410466	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	347107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	126864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	243233	56.384	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.760%
35) Dibromofluoromethane	8.016	113	162467	58.662	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	117.320%
50) Toluene-d8	10.439	98	574199	54.772	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.540%
62) 4-Bromofluorobenzene	12.727	95	183946	46.192	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.380%
Target Compounds						
						Qvalue
80) 1,3,5-Trimethylbenzene	13.057	105	3006	0.330	ug/l	100
84) 1,2,4-Trimethylbenzene	13.362	105	7594	0.870	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
Data File : VN072560.D
Acq On : 19 May 2022 17:15
Operator : JC\MD
Sample : N2965-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30985

Quant Time: May 20 06:54:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
Quant Title : SW846 8260
QLast Update : Thu May 19 02:22:47 2022
Response via : Initial Calibration

