

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
 Data File : VN072563.D
 Acq On : 19 May 2022 18:28
 Operator : JC\MD
 Sample : N2965-05 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31600

Quant Time: May 20 06:55:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 02:22:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	285629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	424487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	371598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	134889	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	245121	55.234	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.460%
35) Dibromofluoromethane	8.022	113	163055	56.929	ug/l	0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	=	113.860%
50) Toluene-d8	10.439	98	586790	54.125	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.240%
62) 4-Bromofluorobenzene	12.727	95	189914	46.116	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
Data File : VN072563.D
Acq On : 19 May 2022 18:28
Operator : JC\MD
Sample : N2965-05 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31600

Quant Time: May 20 06:55:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
Quant Title : SW846 8260
QLast Update : Thu May 19 02:22:47 2022
Response via : Initial Calibration

