

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
 Data File : VN072565.D
 Acq On : 19 May 2022 19:16
 Operator : JC\MD
 Sample : N2965-07 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31608

Quant Time: May 20 06:56:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 02:22:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	286768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	427488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	369581	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	157703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	234432	52.616	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 105.240%		
35) Dibromofluoromethane	8.016	113	155906	54.051	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 108.100%		
50) Toluene-d8	10.439	98	572547	52.440	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 104.880%		
62) 4-Bromofluorobenzene	12.727	95	200737	48.402	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 96.800%		
Target Compounds						
					Qvalue	
22) Diisopropyl ether	6.492	45	10364	1.182	ug/l #	93
31) Cyclohexane	8.092	56	43347	8.753	ug/l #	85
39) Methylcyclohexane	9.457	83	15324	3.344	ug/l	94
40) Benzene	8.457	78	1100136	95.881	ug/l	100
52) Toluene	10.504	92	2873831	386.287	ug/l	99
67) Ethyl Benzene	11.839	91	480790	36.677	ug/l	97
68) m/p-Xylenes	11.945	106	702910	139.280	ug/l	90
69) o-Xylene	12.274	106	329317	64.056	ug/l	90
73) Isopropylbenzene	12.574	105	16880	1.295	ug/l	96
78) n-propylbenzene	12.916	91	44174	3.098	ug/l	99
80) 1,3,5-Trimethylbenzene	13.051	105	65082	5.756	ug/l	94
84) 1,2,4-Trimethylbenzene	13.363	105	286393	26.405	ug/l	95
95) Naphthalene	15.498	128	19399	2.792	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
Data File : VN072565.D
Acq On : 19 May 2022 19:16
Operator : JC\MD
Sample : N2965-07 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31608

Quant Time: May 20 06:56:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
Quant Title : SW846 8260
QLast Update : Thu May 19 02:22:47 2022
Response via : Initial Calibration

