



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051922\
 Data File : VN072548.D
 Acq On : 19 May 2022 11:05
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
 Title : SW846 8260
 Last Update : Thu May 19 02:22:47 2022



AutoFind: Scans 1824, 1825, 1826; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.0	6725	PASS
75	95	30	60	58.2	17814	PASS
95	95	100	100	100.0	30632	PASS
96	95	5	9	7.2	2219	PASS
173	174	0.00	2	0.9	222	PASS
174	95	50	100	78.9	24171	PASS
175	174	5	9	7.6	1845	PASS
176	174	95	101	95.4	23066	PASS
177	176	5	9	7.0	1607	PASS