

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077757.D
 Acq On : 19 May 2023 12:28
 Operator : JC\MD
 Sample : 02834-07DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D1-20230516DL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/22/2023
 Supervised By : Mahesh Dadoda 05/22/2023

Quant Time: May 19 23:54:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	343812	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	620105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	531024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	179319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	233083	45.309	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.620%		
35) Dibromofluoromethane	8.177	113	195608	48.374	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.740%		
50) Toluene-d8	10.577	98	758829	48.755	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.500%		
62) 4-Bromofluorobenzene	12.853	95	248304	45.171	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.340%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.383	101	2114m	0.556	ug/l	Qvalue
27) cis-1,2-Dichloroethene	7.495	96	1562	0.321	ug/l	94
44) Trichloroethene	9.359	130	273646	61.920	ug/l	99
64) Tetrachloroethene	11.106	164	2337	0.629	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077757.D
Acq On : 19 May 2023 12:28
Operator : JC\MD
Sample : 02834-07DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D1-20230516DL

Quant Time: May 19 23:54:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/22/2023
Supervised By :Mahesh Dadoda 05/22/2023

