

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077758.D
 Acq On : 19 May 2023 12:51
 Operator : JC\MD
 Sample : 02834-11DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D1-20230517DL

Quant Time: May 19 23:54:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/22/2023
 Supervised By : Mahesh Dadoda 05/22/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	362183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	656804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	576493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	200245	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	247885	45.742	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.480%		
35) Dibromofluoromethane	8.177	113	206605	48.239	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.480%		
50) Toluene-d8	10.576	98	808269	49.029	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.060%		
62) 4-Bromofluorobenzene	12.859	95	269382	46.267	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.540%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.394	101	3867	0.965	ug/l #	80
27) cis-1,2-Dichloroethene	7.488	96	2619m	0.510	ug/l	
44) Trichloroethene	9.359	130	171797	36.702	ug/l	98
64) Tetrachloroethene	11.112	164	8553	2.122	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077758.D
Acq On : 19 May 2023 12:51
Operator : JC\MD
Sample : 02834-11DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE125D1-20230517DL

Quant Time: May 19 23:54:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/22/2023
Supervised By : Mahesh Dadoda 05/22/2023

