

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077759.D
 Acq On : 19 May 2023 13:15
 Operator : JC\MD
 Sample : 02834-12DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D3-20230517DL

Quant Time: May 19 23:54:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	361690	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	656257	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	579853	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	199961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	246866	45.616	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.240%
35) Dibromofluoromethane	8.171	113	210754	49.249	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.500%
50) Toluene-d8	10.577	98	808996	49.114	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.220%
62) 4-Bromofluorobenzene	12.853	95	262409	45.107	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.220%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.389	101	97741	24.420	ug/l	98
27) cis-1,2-Dichloroethene	7.489	96	1792	0.350	ug/l	95
44) Trichloroethene	9.359	130	110470	23.620	ug/l	95
64) Tetrachloroethene	11.112	164	9085	2.240	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077759.D
Acq On : 19 May 2023 13:15
Operator : JC\MD
Sample : 02834-12DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE125D3-20230517DL

Quant Time: May 19 23:54:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

