

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077760.D
 Acq On : 19 May 2023 13:38
 Operator : JC\MD
 Sample : 02834-13DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D2-20230517DL

Quant Time: May 19 23:54:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	367294	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	658477	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	583043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	209583	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	245351	44.644	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.280%
35) Dibromofluoromethane	8.177	113	210849	49.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.200%
50) Toluene-d8	10.577	98	808617	48.926	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.853	95	269744	46.212	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.420%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.389	101	7727	1.901	ug/l #	81
12) 1,1-Dichloroethene	4.348	96	2459	0.637	ug/l #	72
27) cis-1,2-Dichloroethene	7.494	96	4439	0.853	ug/l	95
44) Trichloroethene	9.359	130	271198	57.790	ug/l	98
64) Tetrachloroethene	11.106	164	7607	1.866	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077760.D
Acq On : 19 May 2023 13:38
Operator : JC\MD
Sample : 02834-13DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE125D2-20230517DL

Quant Time: May 19 23:54:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

