

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077761.D
 Acq On : 19 May 2023 14:02
 Operator : JC\MD
 Sample : 02862-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-TB-2023

Quant Time: May 19 23:54:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	367612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	668324	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	580058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	202458	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	249043	45.277	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.560%		
35) Dibromofluoromethane	8.177	113	209215	48.006	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.020%		
50) Toluene-d8	10.576	98	825138	49.190	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.380%		
62) 4-Bromofluorobenzene	12.853	95	266110	44.917	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.840%		
Target Compounds						
44) Trichloroethene	9.347	130	1773	0.372	ug/l	43

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077761.D
Acq On : 19 May 2023 14:02
Operator : JC\MD
Sample : 02862-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-TB-2023

Quant Time: May 19 23:54:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

