

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077766.D
 Acq On : 19 May 2023 16:00
 Operator : JC\MD
 Sample : 02862-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-58-60

Quant Time: May 19 23:55:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	383157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	680464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.877	117	595374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	203520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	251401	43.851	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.700%		
35) Dibromofluoromethane	8.177	113	216310	48.749	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.500%		
50) Toluene-d8	10.577	98	844017	49.418	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.840%		
62) 4-Bromofluorobenzene	12.859	95	269478	44.674	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.340%		
Target Compounds						
16) Acetone	4.448	43	7450	4.130	ug/l	94
30) Chloroform	7.965	83	7532	0.805	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077766.D
Acq On : 19 May 2023 16:00
Operator : JC\MD
Sample : 02862-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-58-60

Quant Time: May 19 23:55:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

