

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077769.D
 Acq On : 19 May 2023 17:11
 Operator : JC\MD
 Sample : 02862-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-198-200

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/22/2023
 Supervised By : Mahesh Dadoda 05/22/2023

Quant Time: May 19 23:56:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	362554	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	656825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	574653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	198744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	246213	45.387	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.780%		
35) Dibromofluoromethane	8.177	113	207331	48.407	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.820%		
50) Toluene-d8	10.577	98	808532	49.044	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.080%		
62) 4-Bromofluorobenzene	12.853	95	259641	44.593	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.180%		
Target Compounds						
16) Acetone	4.465	43	5773m	3.383	ug/l	
25) 2-Butanone	7.488	43	6078	2.277	ug/l	97
30) Chloroform	7.971	83	2856	0.323	ug/l #	71
44) Trichloroethene	9.371	130	1608	0.344	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077769.D
Acq On : 19 May 2023 17:11
Operator : JC\MD
Sample : 02862-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-198-200

Quant Time: May 19 23:56:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/22/2023
Supervised By :Mahesh Dadoda 05/22/2023

