

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077770.D
 Acq On : 19 May 2023 17:35
 Operator : JC\MD
 Sample : 02862-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-DUP-20230517

Quant Time: May 19 23:56:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	387125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	702541	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	597882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	194343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257903	44.525	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.040%
35) Dibromofluoromethane	8.177	113	220655	48.165	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.340%
50) Toluene-d8	10.576	98	858382	48.679	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.360%
62) 4-Bromofluorobenzene	12.859	95	268914	43.180	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.360%
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	4024	2.208	ug/l	94
30) Chloroform	7.977	83	4244	0.449	ug/l	84
44) Trichloroethene	9.371	130	1708	0.341	ug/l	54
64) Tetrachloroethene	11.112	164	1791	0.428	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077770.D
Acq On : 19 May 2023 17:35
Operator : JC\MD
Sample : 02862-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-DUP-20230517

Quant Time: May 19 23:56:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

