

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077771.D
 Acq On : 19 May 2023 17:58
 Operator : JC\MD
 Sample : 02862-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-218-220

Quant Time: May 19 23:56:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	351426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	636553	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	560584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	192571	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	238877	45.429	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.860%
35) Dibromofluoromethane	8.177	113	201698	48.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.180%
50) Toluene-d8	10.577	98	784734	49.116	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.240%
62) 4-Bromofluorobenzene	12.853	95	253212	44.874	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.740%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	4011	2.425	ug/l #	69
30) Chloroform	7.971	83	3520	0.410	ug/l	98
44) Trichloroethene	9.359	130	2181	0.481	ug/l	63
64) Tetrachloroethene	11.118	164	1649	0.421	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077771.D
Acq On : 19 May 2023 17:58
Operator : JC\MD
Sample : 02862-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-218-220

Quant Time: May 19 23:56:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

