

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077774.D
 Acq On : 19 May 2023 19:09
 Operator : JC\MD
 Sample : 02862-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-278-280

Quant Time: May 19 23:57:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	355908	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	632053	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	561597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	194717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	234967	44.123	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.240%
35) Dibromofluoromethane	8.177	113	203710	49.426	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.860%
50) Toluene-d8	10.571	98	786490	49.577	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.160%
62) 4-Bromofluorobenzene	12.859	95	250738	44.752	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.500%
Target Compounds						
16) Acetone	4.465	43	3744	2.235	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077774.D
Acq On : 19 May 2023 19:09
Operator : JC\MD
Sample : 02862-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-278-280

Quant Time: May 19 23:57:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

