

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
 Data File : VN077775.D
 Acq On : 19 May 2023 19:33
 Operator : JC\MD
 Sample : 02834-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP02-20230517

Quant Time: May 19 23:57:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	351937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	633942	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.877	117	556184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	193853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	242448	46.041	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.080%
35) Dibromofluoromethane	8.177	113	202931	49.090	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	10.577	98	778533	48.929	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.859	95	252442	44.921	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.840%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.389	101	435993	111.948	ug/l	98
12) 1,1-Dichloroethene	4.359	96	4513	1.219	ug/l #	73
27) cis-1,2-Dichloroethene	7.489	96	8234	1.652	ug/l	83
44) Trichloroethene	9.359	130	525561	116.327	ug/l	99
64) Tetrachloroethene	11.112	164	42234	10.858	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051923\
Data File : VN077775.D
Acq On : 19 May 2023 19:33
Operator : JC\MD
Sample : 02834-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP02-20230517

Quant Time: May 19 23:57:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

