

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051925\
 Data File : VN086690.D
 Acq On : 19 May 2025 15:27
 Operator : JC\MD
 Sample : PB168034ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB168034ZHE#07

Quant Time: May 20 03:20:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	327584	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121325	51.684	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.360%			
35) Dibromofluoromethane	8.165	113	101278	51.913	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.820%			
50) Toluene-d8	10.565	98	419554	51.895	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.800%			
62) 4-Bromofluorobenzene	12.847	95	144349	50.566	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.140%			
Target Compounds						
16) Acetone	4.430	43	25580	29.131	ug/l	96
18) Methyl Acetate	5.024	43	4053	1.649	ug/l #	75
20) Methylene Chloride	5.277	84	3314	1.430	ug/l	91
43) Isopropyl Acetate	8.688	43	38960m	6.686	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051925\
Data File : VN086690.D
Acq On : 19 May 2025 15:27
Operator : JC\MD
Sample : PB168034ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB168034ZHE#07

Quant Time: May 20 03:20:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/21/2025
Supervised By :Mahesh Dadoda 05/21/2025

