

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052019\
 Data File : VN055662.D
 Acq On : 20 May 2019 10:00
 Operator : JC/SP
 Sample : VN0520WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0520WBL01

Quant Time: May 20 11:01:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	424785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	606085	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	514349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	184676	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	196216	42.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.28%	
35) Dibromofluoromethane	7.59	113	186180	48.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.72%	
50) Toluene-d8	10.09	98	700028	48.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.18%	
62) 4-Bromofluorobenzene	12.40	95	219468	43.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052019\
Data File : VN055662.D
Acq On : 20 May 2019 10:00
Operator : JC/SP
Sample : VN0520WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0520WBL01

Quant Time: May 20 11:01:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

