





Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052019\  
Data File : VN055664.D  
Acq On : 20 May 2019 11:13  
Operator : JC/SP  
Sample : VN0520WBS01  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0520WBS01

Manual Integrations  
APPROVED

MMDadoda  
5/21/2019 12:38:13 PM

Quant Time: May 21 04:43:20 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N051719W.M  
Quant Title : SW846 8260  
QLast Update : Fri May 17 23:54:12 2019  
Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|-------|------|----------|--------|-------|----------|
| 93) 1,2,4-Trichlorobenzene | 14.91 | 180  | 74869    | 19.596 | ug/l  | 99       |
| 94) Hexachlorobutadiene    | 15.01 | 225  | 44756    | 17.515 | ug/l  | 99       |
| 95) Naphthalene            | 15.13 | 128  | 245105   | 22.271 | ug/l  | 95       |
| 96) 1,2,3-Trichlorobenzene | 15.31 | 180  | 78492    | 19.648 | ug/l  | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052019\  
 Data File : VN055664.D  
 Acq On : 20 May 2019 11:13  
 Operator : JC/SP  
 Sample : VN0520WBS01  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 Client Sample Id :  
 VN0520WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 5/21/2019 12:38:13 PM

Quant Time: May 21 04:43:20 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N051719W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri May 17 23:54:12 2019  
 Response via : Initial Calibration

