

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052019\
 Data File : VN055667.D
 Acq On : 20 May 2019 12:26
 Operator : JC/SP
 Sample : K2945-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OH55-3E-2SB-2CTR2-GY-G1

Quant Time: May 21 04:50:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	449457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	628440	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	529414	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	190077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	201445	40.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.78%	
35) Dibromofluoromethane	7.59	113	191080	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
50) Toluene-d8	10.09	98	724081	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
62) 4-Bromofluorobenzene	12.40	95	225544	43.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.54%	
Target Compounds						
95) Naphthalene	15.13	128	6264	2.376	ug/l #	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052019\
Data File : VN055667.D
Acq On : 20 May 2019 12:26
Operator : JC/SP
Sample : K2945-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OH55-3E-2SB-2CTR2-GY-G1

Quant Time: May 21 04:50:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

