

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052019\
 Data File : VN055668.D
 Acq On : 20 May 2019 12:50
 Operator : JC/SP
 Sample : K2945-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OH55-3E-2SB-2CTR2-GY-G2

Quant Time: May 21 04:51:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	431386	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	609685	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	516429	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	188733	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	195597	41.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.72%	
35) Dibromofluoromethane	7.59	113	185914	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
50) Toluene-d8	10.09	98	708421	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
62) 4-Bromofluorobenzene	12.40	95	219007	43.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.62%	
Target Compounds						
16) Acetone	3.82	43	17226	8.926	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052019\
Data File : VN055668.D
Acq On : 20 May 2019 12:50
Operator : JC/SP
Sample : K2945-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OH55-3E-2SB-2CTR2-GY-G2

Quant Time: May 21 04:51:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

