

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052019\
 Data File : VN055670.D
 Acq On : 20 May 2019 13:39
 Operator : JC/SP
 Sample : K2913-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-17-S

Quant Time: May 21 04:59:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	409412	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	583676	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	491246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	188354	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	186880	41.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.28%	
35) Dibromofluoromethane	7.59	113	176855	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
50) Toluene-d8	10.09	98	669356	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	12.40	95	210351	43.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.90%	
Target Compounds						
16) Acetone	3.82	43	16977	9.269	ug/l	98
20) Methylene Chloride	4.55	84	6902	1.651	ug/l	92
43) Isopropyl Acetate	8.17	43	12917	1.581	ug/l #	92
95) Naphthalene	15.13	128	38927	5.598	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052019\
Data File : VN055670.D
Acq On : 20 May 2019 13:39
Operator : JC/SP
Sample : K2913-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-17-S

Quant Time: May 21 04:59:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

