

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN052019\
 Data File : VN055679.D
 Acq On : 20 May 2019 17:17
 Operator : JC/SP
 Sample : K2894-02
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-04-9-10

Quant Time: May 20 17:39:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	357817	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	508680	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	432249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	178954	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	161770	41.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.48%	
35) Dibromofluoromethane	7.59	113	151698	46.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.90%	
50) Toluene-d8	10.09	98	580455	47.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.02%	
62) 4-Bromofluorobenzene	12.40	95	186535	44.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052019\
Data File : VN055679.D
Acq On : 20 May 2019 17:17
Operator : JC/SP
Sample : K2894-02
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-04-9-10

Quant Time: May 20 17:39:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

