


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN052019\
Data File : VN055659.D
Acq On : 20 May 2019 8:28
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Title : SW846 8260
Last Update : Fri May 17 23:54:12 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.6	7554	PASS
75	95	30	60	45.3	21984	PASS
95	95	100	100	100.0	48520	PASS
96	95	5	9	6.7	3262	PASS
173	174	0.00	2	0.8	382	PASS
174	95	50	100	93.3	45280	PASS
175	174	5	9	7.1	3201	PASS
176	174	95	101	96.0	43458	PASS
177	176	5	9	6.7	2897	PASS