

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
 Data File : VN061506.D
 Acq On : 20 May 2020 17:27
 Operator : JC/MD
 Sample : L2671-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS052MW128-200514

Quant Time: May 21 02:18:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	281815	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	507701	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	491656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	208762	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	186733	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	
35) Dibromofluoromethane	7.55	113	148012	55.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.30%	
50) Toluene-d8	10.06	98	628455	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	12.37	95	235086	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
Target Compounds						
4) Vinyl Chloride	2.16	62	3731	0.682	ug/l	90
27) cis-1,2-Dichloroethene	6.78	96	30839	7.838	ug/l	82
44) Trichloroethene	8.80	130	5065	1.351	ug/l	88
95) Naphthalene	15.11	128	3570	0.324	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052020\
Data File : VN061506.D
Acq On : 20 May 2020 17:27
Operator : JC/MD
Sample : L2671-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS052MW128-200514

Quant Time: May 21 02:18:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

