

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
 Data File : VN061507.D
 Acq On : 20 May 2020 17:50
 Operator : JC/MD
 Sample : L2676-02DL 4X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S32-0689DL

Quant Time: May 21 02:21:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	279637	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	509594	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	489364	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	210124	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	186954	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
35) Dibromofluoromethane	7.55	113	144400	54.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.18%	
50) Toluene-d8	10.06	98	626255	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	12.37	95	234738	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
Target Compounds						
4) Vinyl Chloride	2.16	62	6349	1.170	ug/l #	87
21) trans-1,2-Dichloroethene	4.99	96	25319	7.660	ug/l	91
27) cis-1,2-Dichloroethene	6.78	96	207249	53.083	ug/l	84
44) Trichloroethene	8.80	130	7102	1.888	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052020\
Data File : VN061507.D
Acq On : 20 May 2020 17:50
Operator : JC/MD
Sample : L2676-02DL 4X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S32-0689DL

Quant Time: May 21 02:21:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

