

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
 Data File : VN061511.D
 Acq On : 20 May 2020 19:24
 Operator : JC/MD
 Sample : PB129002TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129002TB

Quant Time: May 21 08:05:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	289296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	529778	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	533709	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	237998	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	197732	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
35) Dibromofluoromethane	7.55	113	136871	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
50) Toluene-d8	10.06	98	668401	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
62) 4-Bromofluorobenzene	12.38	95	259495	53.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.92%	
Target Compounds						
20) Methylene Chloride	4.50	84	11261	3.174	ug/l	93
43) Isopropyl Acetate	8.13	43	80994	10.337	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052020\
Data File : VN061511.D
Acq On : 20 May 2020 19:24
Operator : JC/MD
Sample : PB129002TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129002TB

Quant Time: May 21 08:05:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

