

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
 Data File : VN061512.D
 Acq On : 20 May 2020 19:47
 Operator : JC/MD
 Sample : PB129002ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129002ZHE#01

Quant Time: May 21 02:41:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	270627	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	498854	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	499826	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	221808	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	181516	44.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.88%	
35) Dibromofluoromethane	7.55	113	133773	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
50) Toluene-d8	10.06	98	625695	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
62) 4-Bromofluorobenzene	12.37	95	242052	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
Target Compounds						
20) Methylene Chloride	4.50	84	11529	3.473	ug/l	91
43) Isopropyl Acetate	8.12	43	77262	10.472	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052020\
Data File : VN061512.D
Acq On : 20 May 2020 19:47
Operator : JC/MD
Sample : PB129002ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129002ZHE#01

Quant Time: May 21 02:41:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

