

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
 Data File : VN061513.D
 Acq On : 20 May 2020 20:10
 Operator : JC/MD
 Sample : PB129002ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129002ZHE#02

Quant Time: May 21 02:46:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	279903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	512304	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	516215	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	224185	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	189365	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
35) Dibromofluoromethane	7.55	113	137844	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
50) Toluene-d8	10.06	98	640397	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	12.38	95	250297	53.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.64%	
Target Compounds						
20) Methylene Chloride	4.51	84	11791	3.434	ug/l	92
43) Isopropyl Acetate	8.13	43	76059	10.038	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052020\
Data File : VN061513.D
Acq On : 20 May 2020 20:10
Operator : JC/MD
Sample : PB129002ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129002ZHE#02

Quant Time: May 21 02:46:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

