

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
 Data File : VN061514.D
 Acq On : 20 May 2020 20:34
 Operator : JC/MD
 Sample : PB129002ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129002ZHE#03

Quant Time: May 21 02:53:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	299427	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	552576	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	556069	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	249281	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	203128	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
35) Dibromofluoromethane	7.55	113	140063	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	10.06	98	695636	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
62) 4-Bromofluorobenzene	12.38	95	271370	54.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.20%	
Target Compounds						
20) Methylene Chloride	4.50	84	11692	3.183	ug/l	93
43) Isopropyl Acetate	8.13	43	80714	9.876	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052020\
Data File : VN061514.D
Acq On : 20 May 2020 20:34
Operator : JC/MD
Sample : PB129002ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129002ZHE#03

Quant Time: May 21 02:53:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

