

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
 Data File : VN061515.D
 Acq On : 20 May 2020 20:57
 Operator : JC/MD
 Sample : L2497-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HD-02-051920

Quant Time: May 21 02:57:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	286513	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	522943	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	529558	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	235872	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	193099	45.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.32%	
35) Dibromofluoromethane	7.55	113	150530	54.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.90%	
50) Toluene-d8	10.06	98	658536	51.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.66%	
62) 4-Bromofluorobenzene	12.37	95	256731	54.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.16%	
Target Compounds						
16) Acetone	3.78	43	17846	5.105	ug/l	97
20) Methylene Chloride	4.51	84	14460	4.115	ug/l	94
43) Isopropyl Acetate	8.13	43	71344	9.224	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
Data File : VN061515.D
Acq On : 20 May 2020 20:57
Operator : JC/MD
Sample : L2497-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampled :
HD-02-051920

Quant Time: May 21 02:57:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

