

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052020\
 Data File : VN061516.D
 Acq On : 20 May 2020 21:20
 Operator : JC/MD
 Sample : L2713-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BUS-WASH-SLUDGE-PIT

Quant Time: May 21 03:01:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	268633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	494554	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	494730	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	227575	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	183239	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	
35) Dibromofluoromethane	7.55	113	145430	56.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.26%	
50) Toluene-d8	10.06	98	619795	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
62) 4-Bromofluorobenzene	12.37	95	239618	53.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.74%	
Target Compounds						
16) Acetone	3.77	43	60195	49.512	ug/l	98
20) Methylene Chloride	4.50	84	27431	8.325	ug/l	94
25) 2-Butanone	6.79	43	20129	11.027	ug/l	99
30) Chloroform	7.33	83	9512	1.551	ug/l	94
43) Isopropyl Acetate	8.13	43	64342	8.797	ug/l #	90
51) 4-Methyl-2-Pentanone	9.95	43	67328	16.897	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052020\
Data File : VN061516.D
Acq On : 20 May 2020 21:20
Operator : JC/MD
Sample : L2713-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BUS-WASH-SLUDGE-PIT

Quant Time: May 21 03:01:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

