


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN052020\
 Data File : VN061498.D
 Acq On : 20 May 2020 13:47
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Title : SW846 8260
 Last Update : Mon May 18 13:09:55 2020



AutoFind: Scans 3296, 3297, 3298; Background Corrected with Scan 3283

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	8698	PASS
75	95	30	60	51.1	25200	PASS
95	95	100	100	100.0	49290	PASS
96	95	5	9	7.0	3453	PASS
173	174	0.00	2	1.3	447	PASS
174	95	50	100	71.2	35093	PASS
175	174	5	9	8.0	2792	PASS
176	174	95	101	95.0	33336	PASS
177	176	5	9	7.0	2319	PASS