

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052022\
 Data File : VN072578.D
 Acq On : 20 May 2022 15:36
 Operator : JC\MD
 Sample : N2965-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31508

Quant Time: May 21 01:33:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 02:22:47 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	274792	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	409801	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	357731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	140066	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	238127	55.774	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 111.540%	
35) Dibromofluoromethane	8.022	113	160799	58.154	ug/l	0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	= 116.300%	
50) Toluene-d8	10.439	98	552951	52.831	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.660%	
62) 4-Bromofluorobenzene	12.727	95	187258	47.101	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.200%	
Target Compounds						
						Qvalue
16) Acetone	4.293	43	37311	28.830	ug/l	97
25) 2-Butanone	7.340	43	5672	2.787	ug/l	95
31) Cyclohexane	8.081	56	6903	1.455	ug/l	# 23
52) Toluene	10.504	92	3519	0.493	ug/l	90
67) Ethyl Benzene	11.839	91	9032	0.712	ug/l	# 76
68) m/p-Xylenes	11.945	106	22484	4.603	ug/l	91
69) o-Xylene	12.269	106	3103	0.624	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052022\
Data File : VN072578.D
Acq On : 20 May 2022 15:36
Operator : JC\MD
Sample : N2965-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31508

Quant Time: May 21 01:33:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
Quant Title : SW846 8260
QLast Update : Thu May 19 02:22:47 2022
Response via : Initial Calibration

