



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052022\
 Data File : VN072567.D
 Acq On : 20 May 2022 09:47
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051822W.M
 Title : SW846 8260
 Last Update : Thu May 19 02:22:47 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	6446	PASS
75	95	30	60	58.2	19381	PASS
95	95	100	100	100.0	33288	PASS
96	95	5	9	6.4	2123	PASS
173	174	0.00	2	1.0	285	PASS
174	95	50	100	83.6	27829	PASS
175	174	5	9	7.1	1988	PASS
176	174	95	101	98.8	27493	PASS
177	176	5	9	7.3	2016	PASS